

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com



Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 662017
www.weightron.com



Members of the
UK Weighing Federation
Membership No. 1070

CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of
Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

DENWENT WASTE
UNIT 11,
PHE BRIDGE INDUSTRIAL
PHE BRIDGE
ALFRETON
DENWENT
LOCATION DESSANX
BOTTOM HALL

Machine Details

Cust ref	88196
Make	PNEUMATIC
Model	I 300
Ser No	F805491
Capacity	50000g
e/d	20g / 1
Working cap	

Test Weights

Set No.	Cert No.	Date Cal
1-28	6655	23-11-17

Cert Interval ANNUAL
Next Cert JULY 2019

Tolerance +/-	KL
0-10000	20
10000-40000	40
40000-50000	60

Comments ALL WEIGHTS IN KG
FOUR LIFT - 4660
TEST UNIT - 19360
STOPS JAMMED AT FULL LOAD
ADJUSTED & RE-TESTED

2	4	6
1	3	5

AS FOUND LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
0	0	0	0	0
400	420	+20	420	+20
10000	10080	+80	10080	+80
20000	20100	+100	20100	+100
CAL FAILED TEST STOPPED				

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
0	0	0	0	0
400	400	0	400	0
10000	10000	0	10000	0
20000	20000	0	20000	0
25360	25360	0	25360	0
29360	29360	0	29360	0
35360	35360	0	35360	0
39360	39360	0	39360	0
44020	44020	0	44020	0

Repeatability

Load	Ind	Ind	Ind

Repeatability

Load	Ind	Ind	Ind
44020	44020	44020	44020

Eccentric Load - 6660

Reading	Dev	Reading	Dev
1 7740	+80	2 7740	+80
3 7720	+60	4 7700	+40
5 7700	+40	6 7720	+60
7		8	

Eccentric Load - 6660

Reading	Dev	Reading	Dev
1 6660	0	2 6660	0
3 6660	0	4 6660	0
5 6660	0	6 6660	0
7		8	

Signature

Name CHRIS WILSON

Date 27-7-18

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com



Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 662017
www.weightron.com



Members of the
UK Weighing Federation
Membership No. 1070

CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

DEWENT WASTE LTD
UNIT 11,
PHE BRIDGE INDUSTRIAL
PHE BRIDGE
ALFINGTON
DUNSTON
LOCATION
TOP HAND

Machine Details

Cust ref	88196
Make	WERNER
Model	D400
Ser No	102515
Capacity	50,000g
e/d	20g
Working cap	

Test Weights

Set No.	Cert No.	Date Cal
1-28	6655	23-11-17

Cert Interval	ANNUAL
Next Cert	JUN 2019

Tolerance +/-	KL
0-10000	20
10000-40000	40
40000-50000	60

Comments ALL WEIGHTS IN KL
FORM LIFT - 4660
TEST UNIT - 19360

2	3
1	4

AS FOUND LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
0	0	0	0	0
400	400	0	400	0
10000	10000	0	10000	0
20000	20000	0	20000	0
25360	25360	0	25360	0
29360	29360	0	29360	0
35360	35360	0	35360	0
39360	39360	0	39360	0
44020	44000	-20	44000	-20

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
N/A				

Repeatability

Load	Ind	Ind	Ind
44020	44000	44000	44000

Repeatability

Load	Ind	Ind	Ind
N/A			

Eccentric Load - 6660

Reading	Dev	Reading	Dev		
1	6660	0	2	6660	0
3	6660	0	4	6660	0
5			6		
7			8		

Eccentric Load

Reading	Dev	Reading	Dev
N/A			

Signature C1.3.88.1M

Name CRAIG NEILSON

Date 24-7-18